

THE COMPARISON OF THE PRECISION OF DIFFERENT DENTAL RADIOGRAPHIC METHODS IN MANDIBULAR PERI-IMPLANTARY MEASUREMENTS: AN IN VITRO STUDY

Alt Çene İmplant Çevresi Ölçümlerinde Farklı Dental Radyografik Yöntemlerin Hassasiyetlerinin Karşılaştırılması: In Vitro Çalışma

Hakan Bilhan¹, Onur Geckili¹, Selda Arat Bilhan², Fatih Ayçicek¹, Berkman Albayrak³
 Pelin Bozbulut³, Fatma Ünalın¹

Received: 13/05/2014

Accepted: 13/11/2014

ABSTRACT

Purpose: The objective of this in vitro study was to investigate and compare the precisions of several radiodiagnostic methods used in dentistry for the measurement of peri-implantary sites.

Materials and Methods: Six dental implants were placed in a human cadaver mandible. Periapical radiographs obtained with the parallel as well as the bisecting angle technique, digital and conventional panoramic radiographs were used for implant and peri-implant bone measurements. The measurement results at each implant were statistically analyzed.

Results: The ICC values for the inter-observer reliability were 0.79 for implant diameters and 0.96 for implant lengths. Statistical significance was not detected between the differences of the measurements of the 2 examiners from the original implant dimensions related to anatomic locations. For both of the examiner measurements, significantly less difference from the original implant dimensions was detected in the parallel technique compared to the other techniques ($p < 0.05$).

Conclusion: The present study showed that the most precise peri-implant bone measurements can be obtained from periapical radiographies by using the parallel technique.

Keywords: Panoramic radiography, dental implantation, parallel technique, bisecting technique, periapical radiography

ÖZ

Amaç: Bu in vitro çalışmanın amacı implant çevresi alanların ölçümünde kullanılan çeşitli radyodiagnostik metodların hassasiyetlerini değerlendirmek ve karşılaştırmaktır.

Gereç ve Yöntem: İnsan kadavrasına ait alt çeneye 6 adet dental implant yerleştirildi. İmplant ve implant çevresi kemik ölçümleri için hem paralel hem de açıortay tekniği ile elde edilen periapikal radyografiler, dijital ve konvansiyonel panoramik radyografiler kullanıldı. Her implantın ölçüm sonuçları istatistiksel olarak analiz edildi.

Bulgular: Gözlemciler arası güvenilirlik için sınıf içi korrelasyon değerleri implant çapı için 0.79, implant uzunluğu için 0.96'dır. 2 gözlemcinin ölçümleri ve anatomik konumlara ilişkin implant ebatları arasında istatistiksel olarak anlamlı bir fark bulunamamıştır. Her gözlemcinin ölçümünün, paralel teknik kullanılarak yapılan ölçümlerde diğer tekniklere göre implantların gerçek ebatlarından anlamlı olarak daha az fark gösterdiği belirlenmiştir ($p < 0.05$).

Sonuç: Bu çalışma implant çevresi kemik ölçümünde paralel teknik ile elde edilen periapikal radyografinin en hassas ölçüm tekniği olduğunu göstermiştir.

Anahtar kelimeler: Panoramik radyografi, dental implantasyon, paralel teknik, açıortay tekniği, periapikal radyografi

¹ Department of Prosthodontics Faculty of Dentistry Okan University

² Department of Prosthodontics, Medipol Mega Hospital and University

³ Undergraduate Student, Faculty of Dentistry Istanbul University



Introduction

The measurement of the marginal alveolar bone level changes over time using radiographs has been reported to be an important parameter in the evaluation of dental implant success (1). Thus, radiography plays an important role in clinical routine practice and in research projects evaluating dental implants. It is difficult to determine small changes in bone height clinically, therefore radiography is an important help for detailed measurements. Marginal bone loss can be defined as the distance between the alveolar crest and the cement-enamel border or the border line of rough and polished surfaces in dental implants. Several intra- or extra-oral radiographic methods are used in dentistry. Intra- and extra-oral radiographs can be acquired conventionally as well as digitally. Periapical radiographies acquired by using bisecting angle or parallel techniques are often used in implant dentistry (2-4). Panoramic radiographs are widely used for evaluation of the condition of the bone around implants supporting mandibular overdentures (5-10). Panoramic radiograph includes both the jaws and the teeth and is a simple examination process (11), offering ease of operation and a shorter working time. It should be kept in mind that, especially in patients with severe alveolar resorption, the use of intraoral films may be difficult. On the other hand, there are several studies pointing out the disadvantages of panoramic radiographies in diagnostic imaging performed for detailed evaluation and quantification. Panoramic radiographs have been reported to be unsuitable because of the magnification, distortion, relative unsharpness and superimposition of the cervical spine (12, 13). Many authors only accept the results of measurements obtained from standardized intraoral radiographs. The aim of this in vitro study was to investigate the precision of peri-implantary measurements on different 2-dimensional radiographic methods and help clinicians to decide which technique to use for radiographic measurements.

Materials and Methods

The mandible of a formalin fixed human cadaver (Figure 1) was used for the exposure with different radiographic methods. After removal of the mandible from the cadaver, six implant recipient sites with a safe distance to each other (Figure 2) were prepared following the standard drilling protocol recommended by the manufacturer (Straumann Dental Implants;

Institut Straumann AG, Basel, Switzerland). Care was taken to spread the implants to all possible alveolar sites of the mandible. Six implants were inserted until the roughened surfaces were completely covered by the bone.



Figure 1. The mandible obtained from a formalin fixed human cadaver.



Figure 2. Six implants were placed with a safe distance to each other.

Radiographic Evaluation and Bone Level Assessment

After placement of the implants, following methods were used to obtain radiographies:

1. Conventional panoramic radiography (Planmeca, Proline XC, Helsinki, Finland) (Figure 3).
2. Digital panoramic radiography (Morita Veraview IC5, J. Morita MFG. Corp., Kyoto, Japan).
3. Bisecting angle (Kodak Ultraspeed Dental Film, Eastman Kodak, Rochester, NY) with a Siemens Heliodont MD model X1744 (Sirona Dental Systems, GmbH D-64625, Bensheim, Germany).
4. Parallel technique (Kodak Ultraspeed Dental Film, Eastman Kodak, Rochester, NY) with a Siemens Heliodont MD model X1744 (Sirona Dental Systems, GmbH D-64625, Bensheim, Germany) X-ray machine set to 70 kV and 7 mA.



Figure 3. Conventional panoramic radiography of the mounted mandible.

All the measurements were performed independently by two examiners blinded to the study protocol. Measurements were obtained from images of conventional radiographs which were scanned, digitized (Epson 1680 Pro, Seiko Epson Corporation, Nagano, Japan) and analyzed at 5x magnification using a software program (CorelDraw 11.0; Corel Corp and Coral Ltd, Ottawa, Canada). The implant diameters at the widest region and implant lengths, as specified by the manufacturer, were used as reference points, which were 4.8 mm and 10 mm respectively. To account for variability, implant diameters and lengths were measured and compared to the documented original dimensions, and ratios were calculated to adjust for distortion. A distortion coefficient was determined for each site and imaging technique. Inter-observer reliability was assessed via comparison of the measurements of two different examiners.

Statistical Analysis

Statistical Package for Social Sciences (SPSS) statistical software (ver. 15.0 for Windows; SPSS, Inc., Chicago, IL, USA) was used in this study. Kolmogorov-Smirnov test was used to test normality of the distribution. Since the data did not meet the requirements for normal distribution, non-parametric Kruskal Wallis test was used for multiple comparisons. When significant differences are detected among study variables, Mann Whitney U test was employed for pairwise comparisons. Wilcoxon sign-test was performed to compare the examiners. Inter-observer reliability was calculated by the intra-class correlation

coefficient (ICC). The results were assessed at 95% confidence interval, and p values less than 0.05 are considered significant.

Results

No statistically significant difference was detected between the differences of the overall measurements of the 2 examiners from the original implant lengths and implant diameters ($p > 0.05$; Table 1). The ICC values for the inter-observer reliability were 0.79 for implant diameter and 0.96 for implant lengths. These findings were qualified as excellent agreements according to the recommendations of Schuck (14). The measurements of 2 examiners for in all techniques are presented in Table 2.

No statistical difference was detected between the different measurements of the 2 examiners from the original implant dimensions related to anatomic locations ($p > 0.05$; Table 3).

The measurements of both examiners revealed a significant difference from the original implant dimensions ($p < 0.05$; Table 4). The difference from the original implant dimensions measured by the bisecting angle technique were found significantly lower than the analog and digital panoramic techniques ($p < 0.05$; Table 4). No significance was detected between the differences from the original implant dimensions found in analog and digital panoramic techniques ($p > 0.05$; Table 4).

Discussion

The purpose of this study was to investigate the precision of various radiographic methods in evaluating bone around dental implants. Intraoral and panoramic radiographies were used to compare the precision of measurements in this study. It is noteworthy that the results are reproducible, the ICC values showing excellent inter-observer agreement for overall measurements of two examiners according to the recommendations of Schuck (14).

The peri-implant bone can be analyzed with the aid of vectoral software programs and various measurements on digital or digitized radiographies or by hand tracing of digital radiographic printouts. Two recent articles showed that in computerized tracing of direct digital radiographs compared with hand tracing of digital radiographic printouts, differences were minimal and clinically acceptable (15, 16).

The digital method which was used in the present

study is one of the most often used programs for this purpose. Although a 2000% magnification is recommended, our experience showed that in magnification levels over 500% the image can lose its clarity and the precision of the measurements is lost. For this reason a 500% magnification was used in the present study during measurements. When using radiography techniques, it is essential to use minimal

ionizing radiation doses to obtain images that have acceptable diagnostic quality. Standard methods such as panoramic radiographies or intraoral projections are mostly sufficient, although geometric distortions or magnification in the images may occur depending on the distance of the film to the object as well as position and factors related to the patient.

Table 1. The differences of all of the measurements of the 2 examiners from the original implant dimensions (SD: standard deviation).

		Mean±SD	p
Implant diameter	Examiner 1	0.89±0.72	0.719
	Examiner 2	0.83±0.68	
Implant length	Examiner 1	1.82±1.46	0.586
	Examiner 2	1.85±1.46	

Wilcoxon sign test

Although it had been stated that radiographs cannot always provide diagnostic information, even when a computer-aided analysis system is used (17), radiography plays an important role in clinical routine practice and in research projects evaluating dental implants (18, 19) The bone loss documented

in two-dimensional radiodiagnostic methods is the reduction of the bone levels at the mesial and distal aspects of the implants, ignoring the so-called saucerization of the crestal bone around the neck of the implants since only 2-dimensional imaging is used.

Table 2. The differences of the measurements of the 2 examiners from the original implant dimensions using the bisecting, parallel and analog and digital panoramic techniques.

Technique			Mean±SD	p
Bisecting Technique	Implant Diameter	Examiner 1	0.58±0.46	0.173
		Examiner 2	0.65±0.40	
	Implant length	Examiner 1	1.19±1.93	0.249
		Examiner 2	1.31±1.82	
Parallel Technique	Implant Diameter	Examiner 1	0.05±0.04	0.600
		Examiner 2	0.10±0.15	
	Implant length	Examiner 1	0.13±0.11	0.753
		Examiner 2	0.19±0.30	
Analog panoramic technique	Implant Diameter	Examiner 1	1.67±0.29	0.917
		Examiner 2	1.68±0.33	
	Implant length	Examiner 1	3.40±0.60	0.600
		Examiner 2	3.44±0.68	
Digital panoramic technique	Implant Diameter	Examiner 1	1.42±0.78	0.600
		Examiner 2	1.10±0.77	
	Implant length	Examiner 1	2.90±1.58	0.917
		Examiner 2	3.01±1.39	

Wilcoxon sign test

Table 3. The differences of the measurements of the 2 examiners from the original implant dimensions related to anatomic locations (FDI 2-digit tooth numbering system was used to describe anatomic location, SD: standard deviation).

Anatomic Location	Examiner 1		Examiner 2	
	Implant Diameter	Implant length	Implant Diameter	Implant length
	Mean±SD	Mean±SD	Mean±SD	Mean±SD
Tooth no: 31	0.81±0.63 (0.68)	1.65±1.29 (1.40)	0.77±0.72 (0.71)	1.57±1.48 (1.44)
Tooth no: 33	1.07±0.87 (0.78)	2.22±1.73 (1.60)	1.12±0.94 (0.78)	2.38±2.04 (1.61)
Tooth no: 36	0.41±0.54 (0.05)	0.85±1.10 (0.11)	0.51±0.50 (0.47)	1.04±1.02 (0.97)
Tooth no: 41	1.08±0.74 (0.98)	2.20±1.51 (2.01)	1.06±0.69 (1.10)	2.17±1.42 (2.22)
Tooth no: 43	1.20±0.83 (1.24)	2.46±1.69 (2.54)	0.76±0.71 (0.88)	2.38±1.54 (2.11)
Tooth no: 46	0.76±0.72 (0.53)	1.56±1.47 (1.09)	0.78±0.65 (0.38)	1.60±1.33 (0.78)
p	0.521	0.481	0.856	0.716

Kruskal Wallis Test

Table 4. Comparison of the techniques by subtracting the measured values from the original dimensions (SD: standard deviation).

Techniques	Examiner 1		Examiner 2	
	Implant Diameter	Implant length	Implant Diameter	Implant length
	Mean±SD	Mean±SD	Mean±SD	Mean±SD
Bisecting Technique	0.58±0.46 (0.62)	1.19±1.93 (1.27)	0.65±0.40 (0.67)	1.31±0.82 (1.36)
Parallel Technique	0.05±0.04 (0.04)	0.13±0.11 (0.12)	0.10±0.15 (0.02)	0.19±0.30 (0.06)
Analog panoramic technique	1.66±0.29 (1.7)	3.40±0.60 (3.48)	1.68±0.33 (1.79)	3.44±0.68 (3.67)
Digital panoramic technique	1.42±0.77 (1.61)	2.90±1.58 (3.28)	1.10±0.77 (1.20)	3.01±1.39 (2.96)
p	0.001**	0.001**	0.001**	0.001**

Kruskal Wallis ** $p < 0.01$

Techniques	Examiner 1		Examiner 2	
	Implant Diameter	Implant length	Implant Diameter	Implant length
	p	p	p	p
Bisecting/Parallel Techniques	0.037*	0.037*	0.019*	0.025*
Bisecting/Analog panoramic Techniques	0.006**	0.006**	0.004**	0.004**
Bisecting/Digital panoramic Techniques	0.050*	0.050*	0.262	0.037*
Parallel/Analog panoramic Techniques	0.004**	0.004**	0.004**	0.004**
Parallel/Digital panoramic Techniques	0.011*	0.016*	0.016*	0.004**
Analog/Digital panoramic Techniques	0.631	0.631	0.150	0.631

Mann Whitney U Test * $p < 0.05$, ** $p < 0.01$

To be able to obtain data, especially about the vestibular crestal bone changes, a 3-dimensional imaging technique such as the Cone Beam Computerized Tomography (CBCT) is reported to be necessary (20) and it is considered as the golden standard (2). During radiographic evaluation of a radiolucent zone bordering a metal implant, it should also be reminded that the influence of the Mach-band effect will most probably alter the outcome (21).

Panoramic radiographs are widely used for evaluation of the bone condition around implants (5-10). Although there are numerous studies pointing out the disadvantages of panoramic radiographies in diagnostic imaging (12, 13) and many authorities only accept the results of measurements obtained from standardized intraoral radiographs, it is known that a panoramic radiograph includes both of the jaws and the teeth and is a simple examination method (22). Given that the radiographs were of a high quality, Akesson (23) had concluded that for the assessment of the marginal bone level around teeth the radiographic examination of choice should be the panoramic radiograph, which is in accordance with a study by Persson *et al.* (24). A study by Kullman *et al.* (25) also showed that panoramic radiographs are as reliable as conventional intra-oral films when used to assess the point of bone attachment to implant threads. It was concluded in a recent study that vertical measurements had acceptable accuracy and reproducibility when a software-based calibrated measurement tool was used, confirming that digital panoramic radiography can be reliably utilized to determine the pre-operative implant length in premolar and molar mandibular segments (26). As stated in a recent study, there are only a few studies reporting on the measurement accuracy and on the distortion or magnification of digital panoramic radiography (27-30) and none that describes the impact of measurement accuracy on the pre-operative bone height evaluation prior to especially implant placement in the posterior mandible. Although most panoramic machines also might give varied and unreliable magnification, they are easier to obtain (4), since the panoramic exposure offers ease of operation and shorter working time. In spite of a higher resolution of the images, the most important disadvantage of intraoral radiographies might be that intraoral film-holders can be painful especially for edentulous patients with severe ridge resorption (31, 32). For highly atrophic mandibles with unfavorable imaging conditions, rotational

panoramic radiographs can be a useful alternative to intraoral small-format radiographs for evaluating peri-implant bone loss (32). On conventional panoramic radiographs, dental implants of known length are reported to be easily measured with a sliding caliper to determine the unit's magnification factor (33). Variations in the degree of magnification in the horizontal and vertical planes may result in image distortion, showing a certain degree of vertical magnification due to projection geometry, are expected for conventional film-based panoramic radiography, as well as direct digital panoramic devices (12, 34-36). Due to varying magnification, reference objects with known dimensions are necessary to determine the exact magnification in a particular area (27, 33, 36). Nevertheless, the results of the present study point out the importance of variable distortions in the vertical dimension in measurements on panoramic radiographies. This could be important in implant planning, especially in determination of the implant length, since the planned length can be misleading and implants with undesired dimensions may be used which can lead to injury of anatomic structures such as the mandibular canal or the mental foramen.

According to the results of a recent study (37), the diagnostic performance of conventional and digital panoramic images was reported to be equal for the localization of mental and mandibular foramens. The vertical radiographic measurements were considered generally reliable and had correlated with direct measurements in mandible for conventional and digital panoramic radiography. In the present study the measurements in mandible for conventional and digital panoramic radiography were in correlation as well. Although there are numerous studies showing a sympathy for panoramic radiographies, the results of the present study statistically supported the superiority of the accuracy of measurements on intraoral radiographies and the most superior obtained with the parallel technique.

The comparison of both intraoral techniques has revealed that the parallel method gives much better images suitable for making precise measurements (Figure 4). The main problem with the bisecting angle method was that the placement of the film obviously was inexact in all exposures and the images failed to show the implant threads sharply (Figure 5). Additionally, the bone level at the implant shoulder was not consistent due to exposure geometry. In an experimental study, the errors (38) in radiographic assessment of marginal bone height around

osseointegrated implants had been analyzed and it was demonstrated that very small deviations of the X-ray beam from the perpendicular with respect to the long axis of the implant will compromise the accuracy of bone height measurements. It was pointed out that for taking periapical radiographs the use of the bisecting angle technique should be discouraged because of the inherent distortion of the resultant image. It was explained that the bisecting angle technique relies on a geometric trick to produce the image, but only a portion of the structures being imaged are dimensionally accurate. The long cone paralleling technique for taking periapical radiographs seems to be the technique of choice since it has advantages like reduced skin dose, reduced magnification, showing the true relationship between the bone height and adjacent teeth and the absence of superimposition of the zygoma over the upper molar region. It should be noted that to obtain optimum results from the long cone paralleling technique, the film-focal distance should be approximately 30 cm, as we have used in the present study.




Figure 4. The radiographs taken with the parallel technique give much better images which are suitable for making precise measurements.

Digitizing radiographs were reported to be able to facilitate the quantification of bone changes (39). In a study that had been conducted for the detection of bone loss with different x-ray techniques, it had been reported that a pre-orientation with respect to the expected bone loss is possible using panoramic radiographs (32, 40). Additional intraoral films have reported to be useful where rapid changes of bone levels are expected (40). In the present study, we have used analogue radiographs.




Figure 5. Due to the difficulty of placement of the film in the correct angle images will often fail to show the implant threads sharply.

The daily routine forces us to use more practical methods such as the panoramic radiographies. As an argument it could be stated that the intra-oral radiographic technique can fail to depict the apical portion of the implant. However, this will not be a limitation because in cases where peri-implant bone level changes are evaluated, the region of interest is visualized. If accurate measurements on the radiographies are intended, the parallel technique seems to be the most reliable method to obtain high resolution images to be used for peri-implant bone measurements. Since panoramic and periapical radiographies do not provide three-dimensional information on bone quantity and anatomic structures, CBCT methods are recommended for the measurement of bone resorption around implants (20) and the current literature also inclined towards CBCT analysis (40). This may be regarded as the major limitation of this study and further studies investigating the precision of CBCT methods in the same manner should be conducted.

Conclusion

Within the limitations of this in vitro study, it can be concluded that the most suitable radiographic method for precise mandibular peri-implant bone measurements is achieving the intraoral periapical radiography taken with the parallel technique. This technique is reproducible and provides a good inter-observer reliability.

Source of funding

None declared

Conflict of interest

None declared

References

1. Chaytor DV, Zarb GA, Schmitt A, Lewis DW. The longitudinal effectiveness of osseointegrated dental implants. The Toronto study: bone level changes. *Int J Periodontics Restorative Dent* 1991;11(2):112–125.
2. Monsour PA, Dudhia R. Implant radiography and radiology. *Aust Dent J* 2008;53 (Suppl 1):11-25.
3. Dula K, Mini R, van der Stelt PF, Buser D. The radiographic assessment of implant patients: decision-making criteria. *Int J Oral Maxillofac Implants* 2001;16(1):80–89.
4. Mupparapu M, Singer SR. Implant imaging for the dentist. *J Can Dent Assoc* 2004; 70(1):32.
5. Donatsky O. Osseointegrated dental implants with ball attachments supporting over-dentures in patients with mandibular alveolar ridge atrophy. *Int J Oral Maxillofac Implants* 1993;8(2):162–166.
6. Boerrigter EM, van Oort RP, Raghoobar GM, Stegenga B, Schoen PJ, Boering G. A controlled clinical trial of implant-retained mandibular overdentures: clinical aspects. *J Oral Rehabil* 1997;24(3):182–190.
7. Versteegh PA, Van Beek GJ, Slagter AP, Ottervanger JP. Clinical evaluation of mandibular overdentures supported by multiple-bar fabrication: a follow-up study of two implant systems. *Int J Oral Maxillofac Implants* 1995;10(5):595–603.
8. Mericske-Stern R, Steinlin Schaffner T, Marti P, Geering AH. Peri-implant mucosal aspects of ITI implants supporting overdentures. A five-year longitudinal study. *Clin Oral Implants Res* 1994;5(1):9–18.
9. Spiekermann H, Jansen VK, Richter EJ. A 10-year follow-up study of IMZ and TPS implants in the edentulous mandible using bar-retained overdentures. *Int J Oral Maxillofac Implants* 1995;10(2):231–243.
10. Leimola-Virtanen R, Peltola J, Oksala E, Helenius H, Happonen RP. ITI titanium plasma-sprayed screw implants in the treatment of edentulous mandibles: a follow-up study of 39 patients. *Int J Oral Maxillofac Implants* 1995;10(3):373–378.
11. Akesson L, Hakansson J, Rohlin M, Zöger B. An evaluation of image quality for the assessment of the marginal bone level in panoramic radiography. A comparison of radiographs from different dental clinics. *Swed Dent J* 1993;17(1-2):9–21.
12. Batenburg RHK, Stellingsma K, Raghoobar GM, Vissink A. Bone height measurements on panoramic radiographs: the effect of shape and position of edentulous mandibles. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 1997;84(4): 430–435.
13. Batenburg RH, Meijer HJ, Geraets WG, van der Stelt PF. Radiographic assessment of changes in marginal bone around endosseous implants supporting mandibular overdentures. *Dentomaxillofac Radiol* 1998;27(4):221–224.
14. Schuck P. Assessing reproducibility for interval data in health related quality life questionnaires: which coefficient should be used? *Qual Life Res* 2004;13(3):571–586.
15. Celik E, Polat-Ozsoy O, Toygar Memikoglu TU. Comparison of cephalometric measurements with digital versus conventional cephalometric analysis. *Eur J Orthod* 2009;31(3):241–246.
16. Naoumova J, Lindman R. A comparison of manual traced images and corresponding scanned radiographs digitally traced. *Eur J Orthod* 2009;31(3):247–253.
17. Adell R, Eriksson B, Lekholm U, Brånemark PI, Jemt T. Long-term follow-up study of osseointegrated implants in the treatment of totally edentulous jaws. *Int J Oral Maxillofac Implants* 1990;5(4):347–359.
18. Albrektsson T, Sennerby L. State of the art in oral implants. *J Clin Periodontol* 1991;18(6):474–481.
19. Brägger U, Burgin W, Fourmoussis I, Lang NP. Image processing for the evaluation of dental implants. *Dentomaxillofac Radiol* 1992;21(4):208–212.
20. Duyck J, Rønold HJ, Van Oosterwyck H, Naert I, Vander Sloten J, Ellingsen JE. The influence of static and dynamic loading on marginal bone reactions around osseointegrated implants: an animal experimental study. *Clin Oral Implants Res* 2001;12(3):207–218.
21. Sewerin IP, Gotfredsen K, Stoltze K. Accuracy of radiographic diagnosis of peri-implant radiolucencies--an in vitro experiment. *Clin Oral Implants Res* 1997;8(4):299–304.
22. Misch CE, Perel ML, Wang HL, Sammartino

- G, Galindo-Moreno P, Trisi P, Steigmann M, Rebaudi A, Palti A, Pikos MA, Schwartz-Arad D, Choukroun J, Gutierrez-Perez JL, Marenzi G, Valavanis DK. Implant success, survival, and failure: the International Congress of Oral Implantologists (ICOI) Pisa Consensus Conference. *Implant Dent* 2008;17(1):5-15.
23. Akesson L. Panoramic radiography in the assessment of marginal bone level. *Swed Dent J Suppl* 1991;78:1-129.
 24. Persson RE, Tzannetou S, Feloutzis AG, Brägger U, Persson GR, Lang NP. Comparison between panoramic and intra-oral radiographs for the assessment of alveolar bone levels in a periodontal maintenance population. *J Clin Periodontol* 2003;30(9):833-839.
 25. Kullman L, Al Asfour A, Zetterqvist L, Andersson L. Comparison of radiographic bone height assessments in panoramic and intraoral radiographs of implant patients. *Int J Oral Maxillofac Implants* 2007;22(1):96-100.
 26. Vazquez L, Nizamaldin Y, Combescure C, Nedir R, Bischof M, Dohan Ehrenfest DM, Carrel JP, Belser UC. Accuracy of vertical height measurements on direct digital panoramic radiographs using posterior mandibular implants and metal balls as reference objects. *Dentomaxillofac Radiol* 2013;42(2):20110429.
 27. Schropp L, Stavropoulos A, Gotfredsen E, Wenzel A. Calibration of radiographs by a reference metal ball affects preoperative selection of implant size. *Clin Oral Investig* 2009;13(4):375-381.
 28. Schulze R, Krummenauer F, Schalldach F, d'Hoedt B. Precision and accuracy of measurements in digital panoramic radiography. *Dentomaxillofac Radiol* 2000;29(1): 52-56.
 29. Park JB. The evaluation of digital panoramic radiographs taken for implant dentistry in daily practice. *Med Oral Patol Oral Cir Bucal* 2010;15(4):e663-666.
 30. Kim YK, Park JY, Kim SG, Kim JS, Kim JD. Magnification rate of digital panoramic radiographs and its effectiveness for pre-operative assessment of dental implants. *Dentomaxillofac Radiol* 2011;40(2):76-83.
 31. Chiapasco M. Early and immediate restoration and loading of implants in completely edentulous patients. *Int J Oral Maxillofac Implants* 2004;19 Suppl:76-91.
 32. Zechner W, Watzak G, Gahleitner A, Busenlechner D, Tepper G, Watzek G. Rotational panoramic versus intraoral rectangular radiographs for evaluation of peri-implant bone loss in the anterior atrophic mandible. *Int J Oral Maxillofac Implants* 2003;18(6):873- 878.
 33. Vazquez L, Nizam Al Din Y, Christoph Belser U, Combescure C, Bernard JP. Reliability of the vertical magnification factor on panoramic radiographs: clinical implications for posterior mandibular implants. *Clin Oral Impl Res* 2011;22(12):1420-1425.
 34. Tronje G, Welander U, McDavid WD, Morris CR. Image distortion in rotational panoramic radiography. I. General considerations. *Acta Radiol Diagn (Stockh)* 1981; 22(3A):295-299.
 35. Tronje G, Eliasson S, Julin P, Welander U. Image distortion in rotational panoramic radiography. II. Vertical distances. *Acta Radiol Diagn (Stockh)* 1981;22(4):449-455.
 36. McDavid WD, Dove SB, Welander U, Tronje G. Dimensional reproduction in direct digital rotational panoramic radiography. *Oral Surg Oral Med Oral Pathol* 1993;75(4): 523-527.
 37. Peker I, Gungor K, Semiz M, Tekdemir I. Localization of mental and mandibular foramina on the conventional and digital panoramic images. *Coll Antropol* 2009;33(3):857-862.
 38. Sewerin IP. Errors in radiographic assessment of marginal bone height around osseointegrated implants. *Scand J Dent Res* 1990;98(5):428-433.
 39. Van der Stelt PF. Computer assisted interpretation in radiographic diagnosis. *Dent Clin North Am* 1993;37(4):683-696.
 40. Kim TS, Obst C, Zehaczek S, Geenen C. Detection of bone loss with different X-ray techniques in periodontal patients. *J Periodontol* 2008;79(7):1141-1149.

Corresponding Author:

Onur GECKİLİ

Department of Prosthodontics

Faculty of Dentistry Istanbul University

34093 - Capa, Istanbul, TURKEY

Phone: +90-212-414 20 20 (ext.: 30385)

e-mail: geckili@istanbul.edu.tr